



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/662,001

09/12/2003

John B. Slate

11102.2.1

8607

23862

7590

11/01/2006

NYDEGGER & ASSOCIATES  
348 OLIVE STREET  
SAN DIEGO, CA 92103

EXAMINER

MEHTA, BHISMA

ART UNIT

PAPER NUMBER

3767

DATE MAILED: 11/01/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

<b>Office Action Summary</b>	<b>Application No.</b>	<b>Applicant(s)</b>	
	10/662,001	SLATE ET AL.	
	<b>Examiner</b>	<b>Art Unit</b>	
	Bhisma Mehta	3767	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

#### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

#### Status

- 1) ☒ Responsive to communication(s) filed on 31 August 2006.
- 2a) ☒ This action is **FINAL**.                      2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

#### Disposition of Claims

- 4) ☒ Claim(s) 1-20 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☒ Claim(s) 1-6 is/are allowed.
- 6) ☒ Claim(s) 7 and 10 is/are rejected.
- 7) ☒ Claim(s) 8, 9, and 11-20 is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

#### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

#### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- \* See the attached detailed Office action for a list of the certified copies not received.

#### Attachment(s)

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)                                | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                       | 5) <input type="checkbox"/> Notice of Informal Patent Application                       |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)<br>Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____  |

## **DETAILED ACTION**

### ***Claim Objections***

1. Claims 10-20 are objected to because of the following informalities: Claim 10 recites the limitation ""said injecting step" in lines 1-2. Claim 11 recites the limitations ""said injecting means" and "said infusing step" in line 19. Claim 14 recites the limitation ""said injecting means" in lines 1-2. There is insufficient antecedent basis for these limitations in these claims.

Appropriate correction is required.

### ***Claim Rejections - 35 USC § 102***

2. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

3. Claims 7 and 10 are rejected under 35 U.S.C. 102(b) as being anticipated by Roser (U.S. Patent No. 6,102,896). In line 56 of column 12 to line 12 of column 13, Roser discloses a method of infusing a fluid medicament into a patient which comprises the steps of using a first plunger (34) to inject a partial dose of medicament from an impulse chamber (57) through a nozzle (72) and into a patient at a first fluid pressure and advancing a second plunger (24) to infuse a remainder dose from a reservoir (40) through the nozzle (72) and into a patient at a second fluid pressure. In lines 45-67 of column 13, Roser teaches that the first fluid pressure is greater than the second fluid

Art Unit: 3767

pressure and that the two steps are accomplished sequentially to provide for a substantially continuous, uninterrupted flow of medicament to the patient. As to claim 10, a ram (78) is used for striking the first plunger (34) to advance the first plunger into the impulse chamber.

### ***Response to Arguments***

4. Applicant's arguments, see line 5 of page 13 to line 15 of page 14, filed August 31 2006, with respect to claims 1-10 have been fully considered and are persuasive. The previous prior art rejection of claims 1-10 has been withdrawn.

### ***Allowable Subject Matter***

5. Claims 1-6 are allowed.

6. Claims 8 and 9 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

7. Claims 11-20 would be allowable if rewritten to overcome the objections given above.

### ***Terminal Disclaimer***

8. The terminal disclaimer filed on August 31 2006 disclaiming the terminal portion of any patent granted on this application which would extend beyond the expiration date

Art Unit: 3767

of U.S. Patent No. 6,652,483 has been reviewed and is accepted. The terminal disclaimer has been recorded.

### ***Conclusion***

9. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Bhisma Mehta whose telephone number is 571-272-3383. The examiner can normally be reached on Monday through Friday, 7:30 am to 3:00 pm.

Art Unit: 3767

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Kevin Sirmons can be reached on 571-272-4965. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



BM

KEVIN C. SIRMONS  
SUPERVISORY PATENT EXAMINER

